


Custom-made device

aixstent[®] DUS

Duodenum Stent



a bess group company



bess pro gmbh

Gustav-Krone-Str. 7

D—14167 Berlin

Germany



Leufen Medical GmbH

Gustav-Krone-Str. 7

D—14167 Berlin

Germany

Tel.: +49 30 816 90 93 00

Fax: +49 30 816 90 93 93

www.leufen-medical.eu

contact@leufen-medical.eu

1 Dear Patient,

You have been given an implant of the type aixstent DUS (custom-made device). For your own safety, please read this Patient Information Document carefully and keep it somewhere safe. If you have any questions about your implant, please contact the physician who treats you.

2 About this Document

2.1 Symbols Glossary











Symbol	Description
	MR unsafe
	Catalog number
	Batch code
	Unique Device Identification (UDI)
	Manufacturer
	Distributor
	Patient name
	Date of implantation
	Name of the implanting healthcare institution / provider
	Patient information website

Table 1: Symbols Glossary

2.2 Safety Information Marking

WARNING

Non-compliance may result in serious injuries, serious deterioration of your general condition or your death.

2.3 Additional Information

Download link for the Patient Information Document: ¹⁾	www.leufen-medical.eu/pi/cmlmpi0005
This patient information is based on the following instructions for use:	CMLMGB0005-3

¹⁾ Updated on an ongoing basis.

The catalog number and batch code for your implant can be found on your implant card.

3 What you need to pay attention to

3.1 General

1. Always carry your implant card with you. Show your implant card and this Patient Information Document to your treating physician before undergoing diagnostic or therapeutic procedures.
2. Stick to the appointments you make with your treating physician for follow-up examinations and observe their instructions for any necessary follow-up measures.
3. Contact your doctor if you experience one or more of the following symptoms: Foreign body sensation, pain, bleeding, nausea, vomiting, fever

ATTENTION: Your aixstent DUS must be monitored regularly by your attending physician. Be sure to keep your appointments for these follow-up examinations and follow your physician's advice on the necessary aftercare measures. This is especially true when the intended lifetime of your aixstent DUS has been reached ([▶ Expected Lifetime, page 3]).

3.2 Diet

1. Eat a low-fibre diet.

4 Product Description

4.1 General information

- Self-expanding, woven metal stent
- Without cover / with partial / with complete silicone cover (depending on specifications)
- Atraumatic ends
- Tantalum X-ray markers (depending on specifications)

4.2 Materials with Potential Patient Contact

Product (part)	Material	Contact person	Type of contact
Struts	100% Nitinol ¹⁾	Patient	With every use
Coating (stents with coating only)	100% implant-grade silicone	Patient	With every use
X-ray markers	100% Tantalum	Patient	With every use

¹⁾ Potentially sensitizing / allergenic material

5 Intended Use

5.1 Intended Purpose

The stent is intended to maintain or enable the patency of natural and artificial lumen in the body and/or to cover pathological changes.

5.2 Patient Target Group

The product is suitable for the following patient groups:

- Adults
- Patients of all genders

Custom-made device. The suitability of the product for the patient must be checked and confirmed by the prescribing physician/facility.

5.3 Expected Lifetime

Expected lifetime of the product: 12 months

The product is intended to remain in the body.

6 Possible Complications and Side Effects

The following product-related complications are known:

- Stent breakage
- Bleeding
- Perforations
- Stenosis due to insufficient stent expansion / kinking of the stent
- Stent migration
- Ingrowth of / overgrowth with tissue
- Infection / fever
- Foreign body sensation
- Persistent pain
- Stent occlusion
- Cholangitis
- Cholecystitis
- Pancreatitis
- Liver abscess
- Subcapsular liver hematoma
- Sepsis
- Nausea
- Vomiting

Other known complications such as in endoscopic interventions.

Inter-individual differences in comorbidities and complications may result in some complications becoming more difficult to manage. In rare and extreme cases, this may result in death.

7 Combining with Other Procedures

WARNING

- Laser therapy, argon plasma coagulation, high-frequency surgery, cryotherapy and other procedures, the effect of which is due to heat or cold: Do not use those methods directly on the product. Otherwise, injury to the tissue and product damage are possible.
- MRI safety of the product has not been proven. Therefore, the product must be considered MRI unsafe and must not be used in MR fields.
The possible consequences of the application of non-MRI safe products in MR-fields include: Heating of the product, electromagnetic discharges, consequential damages caused by the application of force to the product, interferences in the imaging (also in the surrounding tissue).

Any method for reducing tissue, such as chemotherapy or radiation therapy, can lead to stent migration.

8 Other Residual Risks

Beyond the listed safety instructions, possible complications and side effects, no further significant residual risks are known.